

EXAMPLES OF SAFETY CRITICAL DEVICES AND ASSOCIATED TEST & INSPECTIONS

Item of Equipment / System	Test/Inspection Required	Suggested Record Keeping <sup>1</sup>
Fume Cabinet alarms and meters: E.g. flow meter, sash height and low flow alarms (also, automatic extinguisher system, if fitted)	Daily Operator Check of flow and warning indicators,	No
	1 - 6 monthly formal check, as appropriate to the risk	Yes

Guards, fixed and movable	Daily Operator Check 6-15.54 <del>8</del> n	No
---------------------------	--	----

Centrifuges, floor standing (e.g. high performance, Ultra)	Daily Inspection	No
	1-2 monthly Operational Test of Emergency power Cut-off(s)	Yes
	1 -2 monthly inspection of Rotors, Interlocks, Cleanliness and Automatic Warning Devices (as applicable)	Yes
	Annual Inspection by External Engineer, more frequent where appropriate	Yes

Lifting Equipment & Lifting Accessories (NON-PEOPLE): E.g. slings, shackles, Fork Lift Truck, MEWPs

Weekly I